

[illegible][illegible][illegible]

<b>4.0</b>	<b>Test gas gas mixture (2.2 % vol. CH<sub>4</sub>, 2.0 % vol. CO<sub>2</sub>, 0 % vol. O<sub>2</sub>, 40 ppm CO, 40 ppm H<sub>2</sub>S)</b>								
4.1	CH <sub>4</sub> - Display 2.00 – 2.40 % vol.								
4.2	CO <sub>2</sub> - Display 1.80 – 2.20 % vol.								
4.3	O <sub>2</sub> - Display -0.5 – +0.5 % vol.								
4.4	CO - Display 37 – 43 ppm								
4.5	H <sub>2</sub> S - Display 37 – 43 ppm								
4.6	Visual alarm (e.g.: Y / N)								
4.7	Audible alarm (e.g.: Y / N)								

[illegible]

<b>6.0 Comments</b> - Maintenance required (inspection plate) - Lifetime sensor exceeded - Adjustment, repair - or similar												
--	--	--	--	--	--	--	--	--	--	--	--	--

[illegible]